




Wall mounted IR Call Point and Ceiling Mounted IR Call Point



The call point can be activated by the push button or infra red trigger. If the infra red trigger is used the identity of the caller and the reset trigger can be identified. The 'call' button will make a patient call pushing both buttons will make a staff call. To reset calls the 'reset' button can be pressed, or an infra red trigger can be used. Another variant has no reset button but has a magnetic swipe reset. The fittings are identical including a semi recessed single gang plate in which alternative labels are affixed.